

資料提供招請に関する公表

次のとおり物品の導入を予定していますので、当該導入に関して資料等の提供を招請します。

令和8年3月4日

独立行政法人国立病院機構嬉野医療センター  
一院長

力武 一久

◎ 調達機関番号 597 ◎ 所在地番号 41

○ 第『13602』号（No.『13602』）

1 調達内容

(1) 品目分類番号 22、31

(2) 導入計画物品及び数量

血管連続撮影装置 1式

(3) 調達方法 購入等

(4) 導入予定時期 令和8年度（要相談）

(5) 調達に必要とされる基本的な要求要件

① 不整脈の診断、治療と頭腹部及び下肢の血管造影による診断、治療および全身の血管造影による診断、治療が安全に行えること

② Cアームは2方向を有し、迅速にアームの角度を設定でき、任意の方向からデジタル透視、デジタル撮影ができること

③ フラットパネルディテクタは、血管造影による診断・治療が行えること

④ デジタル画像処理システムは血管内治療を支援する機能や解析機能を有すること

#### 1 資料及びコメントの提供方法

上記 1 (2) の物品に関する一般的な参考資料及び同 (5) の要求要件等に関するコメント並びに提供可能なライブラリーに関する資料等の提供を招請する。

(1) 資料等の提供期限 令和 8 年 4 月 20 日 17 時 00 分 ( 郵送の場合は必着のこと。 )

(2) 提出先 〒 843-0393 佐賀県嬉野市嬉野町大字下宿甲 4760-1 独立行政法人国立病院機構嬉野医療センター 企画課長 佐野 浩史  
0954-43-1120 内線 220

3 説明書の交付 本公表に基づき応募する供給者に対して導入説明書を交付する。

(1) 交付期間 令和 8 年 3 月 4 日から令和 8 年 4 月 20 日まで。

(2) 交付場所 上記 2 (2) に同じ。

#### 4 その他

この導入計画の詳細は導入説明書による。

なお、本公表内容は予定であり、変更することがあり得る。

## 5 Summary

(1) Classification of the services to be procured: 22,31

(2) Nature and quantity of the services to be required: Serial Angiographic System 1 set

(3) Type of the procurement: Purchase

(4) Introduction schedule time: Within fiscal year 2026

(5) Basic requirements of the procurement :  
Serial Angiographic System

① The system should be designed safely for interventional procedures in arrhythmia, also in head, abdomen, lower extremities and whole body for interventional angiography.

② C-arm should equip either two direction, and the angle of it should enable to set rapidly and simultaneously, in addition, should be capable of digital imaging and fluoroscopy at arbitrary direction.

③ The flat-panel detector should be esp

pecially designed to fulfill the interve  
ntional angiography.

④ The digital imaging processing system  
should have supporting and analysis fun  
ction for intravascular treatment.

(6) Time-limit for the submission of the  
requested material : 5:00P.M. April 20,  
2026

(7) Contact point for the notice: Koji Sano,  
Director, Accounting Divishion, National Hos  
pital Organization Uresino Medical Center,  
4760-1 0-aza Shimojyukukou Ureshino-cho Ur  
eshino-shi Saga-ken 843-0393, Japan. TEL 09  
54-43-1120 ext. 220